

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089927.D **Time Analyzed:** 18:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.069	8.971	9.171	42.760	50.000	-14.5
Endrin	6.572	6.472	6.672	44.560	50.000	-10.9
gamma-BHC (Lindane)	4.328	4.228	4.428	46.890	50.000	-6.2
Heptachlor	4.927	4.827	5.027	45.460	50.000	-9.1
Heptachlor epoxide	5.688	5.589	5.789	46.090	50.000	-7.8
Methoxychlor	7.491	7.392	7.592	43.140	50.000	-13.7
Tetrachloro-m-xylene	3.548	3.448	3.648	45.800	50.000	-8.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089927.D **Time Analyzed:** 18:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.066	7.968	8.168	49.440	50.000	-1.1
Endrin	5.785	5.685	5.885	58.720	50.000	17.4
gamma-BHC (Lindane)	3.727	3.626	3.826	55.750	50.000	11.5
Heptachlor	4.079	3.979	4.179	54.290	50.000	8.6
Heptachlor epoxide	4.869	4.769	4.969	54.870	50.000	9.7
Methoxychlor	6.749	6.650	6.850	56.650	50.000	13.3
Tetrachloro-m-xylene	2.878	2.778	2.978	55.020	50.000	10.0