

Surrogate Summary

SW-846

SDG No.: Q2832

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169228BL	PB169228BL	2-Fluorophenol	150	124	82		30 (18)	130 (112)
		Phenol-d6	150	118	79		30 (15)	130 (107)
		Nitrobenzene-d5	100	76.2	76		30 (18)	130 (107)
		2-Fluorobiphenyl	100	75.0	75		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	124	83		30 (10)	130 (116)
		Terphenyl-d14	100	73.2	73		30 (10)	130 (105)
PB169228BS	PB169228BS	2-Fluorophenol	150	125	84		30 (18)	130 (112)
		Phenol-d6	150	126	84		30 (15)	130 (107)
		Nitrobenzene-d5	100	76.0	76		30 (18)	130 (107)
		2-Fluorobiphenyl	100	73.2	73		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	128	85		30 (10)	130 (116)
		Terphenyl-d14	100	77.4	77		30 (10)	130 (105)
Q2820-17MS	88H-EMS	2-Fluorophenol	150	84.6	56		30 (18)	130 (112)
		Phenol-d6	150	85.7	57		30 (15)	130 (107)
		Nitrobenzene-d5	100	51.9	52		30 (18)	130 (107)
		2-Fluorobiphenyl	100	49.8	50		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	84.8	57		30 (10)	130 (116)
		Terphenyl-d14	100	45.7	46		30 (10)	130 (105)
Q2820-17MSD	88H-EMSD	2-Fluorophenol	150	88.1	59		30 (18)	130 (112)
		Phenol-d6	150	89.6	60		30 (15)	130 (107)
		Nitrobenzene-d5	100	54.5	55		30 (18)	130 (107)
		2-Fluorobiphenyl	100	50.8	51		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	89.7	60		30 (10)	130 (116)
		Terphenyl-d14	100	48.8	49		30 (10)	130 (105)
Q2832-01	TG-S01	2-Fluorophenol	150	69.3	46		30 (18)	130 (112)
		Phenol-d6	150	73.1	49		30 (15)	130 (107)
		Nitrobenzene-d5	100	43.4	43		30 (18)	130 (107)
		2-Fluorobiphenyl	100	47.0	47		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	72.0	48		30 (10)	130 (116)
		Terphenyl-d14	100	50.2	50		30 (10)	130 (105)
Q2832-03	TG-S02	2-Fluorophenol	150	67.0	45		30 (18)	130 (112)
		Phenol-d6	150	67.8	45		30 (15)	130 (107)
		Nitrobenzene-d5	100	41.6	42		30 (18)	130 (107)
		2-Fluorobiphenyl	100	43.2	43		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	64.3	43		30 (10)	130 (116)
		Terphenyl-d14	100	47.0	47		30 (10)	130 (105)
Q2832-05	TG-S03	2-Fluorophenol	150	101	67		30 (18)	130 (112)
		Phenol-d6	150	101	68		30 (15)	130 (107)
		Nitrobenzene-d5	100	65.6	66		30 (18)	130 (107)
		2-Fluorobiphenyl	100	67.6	68		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	101	67		30 (10)	130 (116)
		Terphenyl-d14	100	73.8	74		30 (10)	130 (105)
Q2832-07	TG-S04	2-Fluorophenol	150	95.5	64		30 (18)	130 (112)
		Phenol-d6	150	96.0	64		30 (15)	130 (107)
		Nitrobenzene-d5	100	60.7	61		30 (18)	130 (107)
		2-Fluorobiphenyl	100	63.8	64		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	94.0	63		30 (10)	130 (116)
		Terphenyl-d14	100	64.6	65		30 (10)	130 (105)
Q2832-09	TG-S05	2-Fluorophenol	150	92.6	62		30 (18)	130 (112)
		Phenol-d6	150	97.2	65		30 (15)	130 (107)
		Nitrobenzene-d5	100	73.2	73		30 (18)	130 (107)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SW-846

SDG No.: Q2832

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2832-09	TG-S05	2-Fluorobiphenyl	100	77.0	77		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	88.8	59		30 (10)	130 (116)
		Terphenyl-d14	100	74.5	74		30 (10)	130 (105)