

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2832

Client ID : SSTDCCC040

Date Analyzed: 08/18/2025

Lab File ID: BP025455.D

Time Analyzed: 10:39

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	129439	7.778	550485	10.55	359982	14.40
UPPER LIMIT	258878	8.278	1100970	11.048	719964	14.901
LOWER LIMIT	64719.5	7.278	275243	10.048	179991	13.901
EPA SAMPLE NO.						
01 TG-S03	166743	7.79	660243	10.56	424895	14.41
02 TG-S04	164690	7.79	657475	10.56	417913	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.