

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2832

Client ID : SSTDCCC040

Date Analyzed: 08/21/2025

Lab File ID: BP025528.D

Time Analyzed: 06:42

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	210336	7.79	893021	10.56	582847	14.40
UPPER LIMIT	420672	8.29	1786040	11.06	1165690	14.901
LOWER LIMIT	105168	7.29	446511	10.06	291424	13.901
EPA SAMPLE NO.						
01 TG-S05	185117	7.78	725235	10.55	450660	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.