

Surrogate Summary

SW-846

SDG No.: Q2832

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169212TB	PB169212TB	2-Fluorophenol	150	124	83		15 (23)	110 (138)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	89.6	90		30 (67)	130 (132)
		2-Fluorobiphenyl	100	84.4	84		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	135	90		15 (44)	110 (137)
		Terphenyl-d14	100	88.3	88		30 (42)	130 (152)
PB169275BL	PB169275BL	2-Fluorophenol	150	122	81		15 (23)	110 (138)
		Phenol-d6	150	120	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	74.3	74		30 (67)	130 (132)
		2-Fluorobiphenyl	100	72.4	72		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	119	79		15 (44)	110 (137)
		Terphenyl-d14	100	76.2	76		30 (42)	130 (152)
PB169275BS	PB169275BS	2-Fluorophenol	150	115	77		15 (23)	110 (138)
		Phenol-d6	150	112	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	71.9	72		30 (67)	130 (132)
		2-Fluorobiphenyl	100	70.6	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	108	72		15 (44)	110 (137)
		Terphenyl-d14	100	68.6	69		30 (42)	130 (152)
Q2832-02	TG-S01	2-Fluorophenol	150	117	78		15 (23)	110 (138)
		Phenol-d6	150	109	72		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.9	79		30 (67)	130 (132)
		2-Fluorobiphenyl	100	74.3	74		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	127	85		15 (44)	110 (137)
		Terphenyl-d14	100	75.6	76		30 (42)	130 (152)
Q2832-02MS	TG-S01MS	2-Fluorophenol	150	109	73		15 (23)	110 (138)
		Phenol-d6	150	104	69		15 (10)	110 (134)
		Nitrobenzene-d5	100	73.3	73		30 (67)	130 (132)
		2-Fluorobiphenyl	100	67.2	67		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	114	76		15 (44)	110 (137)
		Terphenyl-d14	100	73.2	73		30 (42)	130 (152)
Q2832-02MSD	TG-S01MSD	2-Fluorophenol	150	122	81		15 (23)	110 (138)
		Phenol-d6	150	118	78		15 (10)	110 (134)
		Nitrobenzene-d5	100	80.8	81		30 (67)	130 (132)
		2-Fluorobiphenyl	100	74.9	75		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	83.6	84		30 (42)	130 (152)
Q2832-04	TG-S02	2-Fluorophenol	150	127	85		15 (23)	110 (138)
		Phenol-d6	150	118	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	86.5	86		30 (67)	130 (132)
		2-Fluorobiphenyl	100	79.1	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	143	95		15 (44)	110 (137)
		Terphenyl-d14	100	83.1	83		30 (42)	130 (152)
Q2832-06	TG-S03	2-Fluorophenol	150	115	76		15 (23)	110 (138)
		Phenol-d6	150	106	71		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.7	78		30 (67)	130 (132)
		2-Fluorobiphenyl	100	73.9	74		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	126	84		15 (44)	110 (137)
		Terphenyl-d14	100	76.9	77		30 (42)	130 (152)
Q2832-08	TG-S04	2-Fluorophenol	150	115	77		15 (23)	110 (138)
		Phenol-d6	150	112	75		15 (10)	110 (134)
		Nitrobenzene-d5	100	76.1	76		30 (67)	130 (132)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SW-846

SDG No.: Q2832

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2832-08	TG-S04	2-Fluorobiphenyl	100	71.4	71		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	125	83		15 (44)	110 (137)
		Terphenyl-d14	100	75.7	76		30 (42)	130 (152)
Q2832-10	TG-S05	2-Fluorophenol	150	131	87		15 (23)	110 (138)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	102	102		30 (67)	130 (132)
		2-Fluorobiphenyl	100	94.4	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	157	105		15 (44)	110 (137)
		Terphenyl-d14	100	96.0	96		30 (42)	130 (152)