

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PORT06
 Lab Code: ACE SDG NO.: Q2832
 Lab File ID: VN087525.D Date Analyzed: 08/13/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	297883	8.21	560249	9.09	506486	11.85
UPPER LIMIT	595766	8.706	1120500	9.588	1012970	12.347
LOWER LIMIT	148942	7.706	280125	8.588	253243	11.347
EPA SAMPLE NO.						
TG-S03	199451	8.21	417141	9.09	385827	11.85
TG-S04	164010	8.21	350645	9.08	322702	11.85
TG-S05	170102	8.21	344303	9.08	313708	11.85
VN0813WBL01	240605	8.21	517097	9.09	472576	11.85
VN0813WBS01	271136	8.21	532074	9.09	483502	11.85
VN0813WBSD01	266779	8.21	512412	9.08	464252	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.