

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>PORT06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2832</u>
Lab File ID:	<u>VN087525.D</u>	Date Analyzed:	<u>08/13/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>10:57</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	261423	13.77				
UPPER LIMIT	522846	14.27				
LOWER LIMIT	130712	13.27				
EPA SAMPLE NO.						
TG-S03	173385	13.77				
TG-S04	146190	13.77				
TG-S05	145236	13.77				
VN0813WBL01	217582	13.77				
VN0813WBS01	246357	13.77				
VN0813WBSD01	233575	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.