

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PORT06
 Lab Code: ACE SDG NO.: Q2832
 Lab File ID: VN087552.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	255173	13.77				
UPPER LIMIT	510346	14.27				
LOWER LIMIT	127587	13.27				
EPA SAMPLE NO.						
TG-S01	224457	13.77				
TG-S02	236126	13.77				
VN0814WBL01	234765	13.77				
VN0814WBS01	221983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.