

Surrogate Summary

SDG No.: **Q2832**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2832-01	TG-S01	1,2-Dichloroethane-d4	50	56.9	114		70 (63)	130 (155)
		Dibromofluoromethane	50	51.0	102		70 (70)	130 (134)
		Toluene-d8	50	51.8	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.1	96		70 (17)	130 (146)
Q2832-03	TG-S02	1,2-Dichloroethane-d4	50	59.2	118		70 (63)	130 (155)
		Dibromofluoromethane	50	52.1	104		70 (70)	130 (134)
		Toluene-d8	50	51.9	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.5	95		70 (17)	130 (146)
Q2832-05	TG-S03	1,2-Dichloroethane-d4	50	59.2	118		70 (63)	130 (155)
		Dibromofluoromethane	50	51.4	103		70 (70)	130 (134)
		Toluene-d8	50	52.6	105		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.8	98		70 (17)	130 (146)
Q2832-07	TG-S04	1,2-Dichloroethane-d4	50	59.2	118		70 (63)	130 (155)
		Dibromofluoromethane	50	52.8	106		70 (70)	130 (134)
		Toluene-d8	50	51.6	103		70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.7	95		70 (17)	130 (146)
Q2832-09	TG-S05	1,2-Dichloroethane-d4	50	57.6	115		70 (63)	130 (155)
		Dibromofluoromethane	50	52.8	106		70 (70)	130 (134)
		Toluene-d8	50	51.0	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	39.9	80		70 (17)	130 (146)
VY0815SBL01	VY0815SBL01	1,2-Dichloroethane-d4	50	59.8	120		70 (63)	130 (155)
		Dibromofluoromethane	50	52.7	105		70 (70)	130 (134)
		Toluene-d8	50	52.4	105		70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.5	95		70 (17)	130 (146)
VY0815SBS01	VY0815SBS01	1,2-Dichloroethane-d4	50	51.7	103		70 (63)	130 (155)
		Dibromofluoromethane	50	49.3	99		70 (70)	130 (134)
		Toluene-d8	50	50.2	100		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.5	97		70 (17)	130 (146)
VY0815SBSD01	VY0815SBSD01	1,2-Dichloroethane-d4	50	52.1	104		70 (63)	130 (155)
		Dibromofluoromethane	50	50.3	101		70 (70)	130 (134)
		Toluene-d8	50	50.5	101		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.7	97		70 (17)	130 (146)

