

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PORT06
 Lab Code: ACE SDG NO.: Q2832
 Lab File ID: VY023111.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	443799	7.71	696916	8.62	607348	11.41
UPPER LIMIT	887598	8.207	1393830	9.116	1214700	11.914
LOWER LIMIT	221900	7.207	348458	8.116	303674	10.914
EPA SAMPLE NO.						
TG-S01	579878	7.71	1082547	8.61	969395	11.41
TG-S02	570487	7.71	1076714	8.62	962307	11.41
TG-S03	553899	7.71	1043284	8.61	947118	11.41
TG-S04	567197	7.71	1071652	8.61	957867	11.41
TG-S05	551145	7.71	1030444	8.62	844477	11.41
VY0815SBL01	550500	7.71	1049087	8.62	953139	11.41
VY0815SBS01	452403	7.71	727202	8.62	654458	11.41
VY0815SBSD01	440097	7.71	718363	8.62	623331	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.