

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PORT06
 Lab Code: ACE SDG NO.: Q2832
 Lab File ID: VY023111.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	306948	13.347				
UPPER LIMIT	613896	13.847				
LOWER LIMIT	153474	12.847				
EPA SAMPLE NO.						
TG-S01	391179	13.34				
TG-S02	380373	13.34				
TG-S03	378902	13.34				
TG-S04	386816	13.34				
TG-S05	274172	13.34				
VY0815SBL01	370974	13.35				
VY0815SBS01	327154	13.34				
VY0815SBSD01	315352	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.