

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Lab File ID: VY023048.D

BFB Injection Date: 08/12/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:26

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.2
75	30.0 - 60.0% of mass 95	51.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	83.2
175	5.0 - 9.0% of mass 174	6.3 (7.6) 1
176	95.0 - 101.0% of mass 174	80.9 (97.2) 1
177	5.0 - 9.0% of mass 176	5.1 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY023050.D	08/12/2025	14:54
VSTDIC010	VSTDIC010	VY023051.D	08/12/2025	15:17
VSTDIC020	VSTDIC020	VY023052.D	08/12/2025	15:40
VSTDIC050	VSTDIC050	VY023053.D	08/12/2025	16:02
VSTDIC100	VSTDIC100	VY023054.D	08/12/2025	16:25
VSTDIC150	VSTDIC150	VY023055.D	08/12/2025	16:47