

Initial and Continuing Calibration Verification

Client: Scientific Design

SDG No.: Q2835

Project: Central Wastewater Treatment Facility 2025

RunNo.: LB136825

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	48.931	50	98	95-105	08/15/2025
Sample ID: CCV2 COD	mg/L	48.931	50	98	95-105	08/15/2025
Sample ID: CCV3 COD	mg/L	49.946	50	100	95-105	08/15/2025

Initial and Continuing Calibration Verification

Client: Scientific Design

SDG No.: Q2835

Project: Central Wastewater Treatment Facility 2025

RunNo.: LB136875

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.02	7	100	90-110	08/12/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7	100	90-110	08/12/2025
Sample ID:	CCV2						
Field pH		pH	7.00	7	100	90-110	08/12/2025