

### Matrix Spike Summary

<b>Client:</b>	Scientific Design	<b>SDG No.:</b>	Q2835
<b>Project:</b>	Central Wastewater Treatment Facility 2025	<b>Sample ID:</b>	Q2862-05
<b>Client ID:</b>	DSN003MS	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	80.4		29.6		50.0	1	102		08/15/2025

### Matrix Spike Summary

<b>Client:</b>	Scientific Design	<b>SDG No.:</b>	Q2835
<b>Project:</b>	Central Wastewater Treatment Facility 2025	<b>Sample ID:</b>	Q2862-05
<b>Client ID:</b>	DSN003MSD	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	78.4		29.6		50.0	1	98		08/15/2025