

Duplicate Sample Summary

Client: Scientific Design	SDG No.: Q2835
Project: Central Wastewater Treatment Facility 2025	Sample ID: Q2835-01
Client ID: ManufacturingDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	7.18		7.17		1	0.14		08/12/2025

Duplicate Sample Summary

Client: Scientific Design	SDG No.: Q2835
Project: Central Wastewater Treatment Facility 2025	Sample ID: Q2835-02
Client ID: Pilot PlantDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	1.00	U	1.00	U	1	0		08/15/2025
Field pH	pH	+/-20	6.52		6.53		1	0.15		08/12/2025

Duplicate Sample Summary

Client:	Scientific Design	SDG No.:	Q2835
Project:	Central Wastewater Treatment Facility 2025	Sample ID:	Q2848-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	150		140		1	6.9		08/14/2025

Duplicate Sample Summary

Client:	Scientific Design	SDG No.:	Q2835
Project:	Central Wastewater Treatment Facility 2025	Sample ID:	Q2862-05
Client ID:	DSN003DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	29.6		29.6		1	0		08/15/2025

Duplicate Sample Summary

Client:	Scientific Design	SDG No.:	Q2835
Project:	Central Wastewater Treatment Facility 2025	Sample ID:	Q2862-05
Client ID:	DSN003MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	80.4		78.4		1	2.52		08/15/2025