

CASE NARRATIVE

ENTACT

Project Name: 540 Degraw St, Brooklyn, NY - E9309

Project # N/A

Order ID # Q2836

Test Name: TCLP Herbicide

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 08/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TCLP Herbicide. This data package contains results for TCLP Herbicide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324. The analysis of TCLP Herbicides was based on method 8151A and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for TG-S01MS [2,4-DCAA(1)155%, 2,4-DCAA(2)175%], TG-S01MSD [2,4-DCAA(1)156%, 2,4-DCAA(2)176%]-These compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference and these failures confirmed with their original sample. And for WC-A5-02-C [2,4-DCAA(2)133%] This compound did not meet the NJDKQP criteria but met the in-house criteria.

The Retention Times were met for all analysis.

The MS {Q2832-02MS} with File ID: PS031462.D recoveries met the requirements for all compounds except for [2,4,5-TP(Silvex)(1)151% - 2,4,5-TP(Silvex)(2)152%] and [2,4-D(1)183% - 2,4-D(2)166%] These compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The MSD {Q2832-02MSD} with File ID: PS031463.D recoveries met the requirements for all compounds except for [2,4,5-TP(Silvex)(1)153% - 2,4,5-TP(Silvex)(2)161%] and [2,4-D(1)184% - 2,4-D(2)159%] These compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

As per special requirement for this project form-1 are reported in mg/l.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____