



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136792

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	08/12/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	08/12/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	08/12/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	08/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136793

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	08/12/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	08/12/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	08/12/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	08/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136800

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	08/13/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	08/13/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	08/13/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	08/13/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136806

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 ASTM COD	mg/L	49.946	50	100	95-105	08/14/2025
Sample ID: CCV2 ASTM COD	mg/L	50.962	50	102	95-105	08/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136826

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	1	1	100	90-110	08/14/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	0.96	1	96	90-110	08/14/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	0.97	1	97	90-110	08/14/2025
Sample ID: CCV3 ASTM Ammonia	mg/L	1	1	100	90-110	08/14/2025