

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	08/13/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136806

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID: CCB1 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	08/14/2025
Sample ID: CCB2 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	08/14/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136826

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/14/2025
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/14/2025
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/14/2025
Sample ID: CCB3 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/14/2025