

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136806

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136806BS							
ASTM COD	mg/L	50	48.9		98	1	90-110	08/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136812

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136812BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	08/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136853

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136853BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	08/18/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2836

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136826

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169237BS							
ASTM Ammonia	mg/L	1	0.99		99	1	90-110	08/14/2025