

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2836</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2838-08</u>	Client ID:	<u>TP-10DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2838-08DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1070		1550		37	*	P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Copper	ug/L	20	100	U	27.6	J	200.0		P
Lead	ug/L	20	13.3	J	23.1	J	54		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	23.2	J	32.9	J	35		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	344		489		35		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2836</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2838-08MS</u>	Client ID:	<u>TP-10MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2838-08MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4060		4030		1		P
Barium	ug/L	20	2660		2490		7		P
Cadmium	ug/L	20	962		949		1		P
Chromium	ug/L	20	1970		1960		1		P
Copper	ug/L	20	1460		1450		1		P
Lead	ug/L	20	4520		4460		1		P
Mercury	ug/L	20	39.4		35.8		10		CV
Nickel	ug/L	20	2460		2420		2		P
Selenium	ug/L	20	9990		9980		0		P
Silver	ug/L	20	380		371		2		P
Zinc	ug/L	20	1570		1510		4		P