

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2836

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169234BS Mercury	ug/L	4.0	3.94		98	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT
Contract: ENTA05

SDG No.: Q2836
Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169236BS							
Arsenic	ug/L	4000	4010		100	80 - 120	P
Barium	ug/L	1000	982		98	80 - 120	P
Cadmium	ug/L	1000	972		97	80 - 120	P
Chromium	ug/L	2000	2110		106	80 - 120	P
Copper	ug/L	1500	1580		105	80 - 120	P
Lead	ug/L	5000	4900		98	80 - 120	P
Nickel	ug/L	2500	2510		100	80 - 120	P
Selenium	ug/L	10000	10200		102	80 - 120	P
Silver	ug/L	380	390		103	80 - 120	P
Zinc	ug/L	1000	1060		106	80 - 120	P