

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ENTA05

Lab Code: ACE SDG NO.: Q2836

Instrument ID: ECD_P Date(s) Analyzed: 08/01/2025 08/01/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.86	4.76	4.96	16763600
		2	4.94	4.84	5.04	12601100
		3	5.02	4.92	5.12	36635200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.02	4.92	5.12	29037400
		2	5.55	5.45	5.65	15042700
		3	5.83	5.73	5.93	29001800
		4	5.99	5.89	6.09	15468800
		5	6.08	5.98	6.18	11716000
Aroclor-1248	500	1	5.81	5.71	5.91	28382600
		2	6.08	5.98	6.18	40187600
		3	6.28	6.18	6.38	44640800
		4	6.68	6.58	6.78	51582200
		5	6.72	6.62	6.82	53240200
Aroclor-1262	500	1	8.24	8.14	8.34	75789600
		2	8.56	8.46	8.66	130430000
		3	8.89	8.79	8.99	94070800
		4	8.97	8.87	9.07	72851600
		5	9.65	9.55	9.75	50922600
Aroclor-1268	500	1	8.88	8.78	8.98	158028000
		2	8.98	8.88	9.08	146028000
		3	9.22	9.12	9.32	121402000
		4	9.64	9.54	9.74	59648200
		5	10.08	9.98	10.18	357940000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ENTA05

Lab Code: ACE SDG NO.: Q2836

Instrument ID: ECD_P Date(s) Analyzed: 08/01/2025 08/01/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	49085200
		2	4.10	4.00	4.20	35889400
		3	4.18	4.08	4.28	135225000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.18	4.08	4.28	101362000
		2	4.90	4.80	5.00	174266000
		3	5.08	4.98	5.18	45816600
		4	5.16	5.06	5.26	58010800
		5	5.34	5.24	5.44	46979400
Aroclor-1248	500	1	4.90	4.80	5.00	223628000
		2	5.12	5.02	5.22	147770000
		3	5.16	5.06	5.26	179546000
		4	5.34	5.24	5.44	167057000
		5	5.73	5.63	5.83	302782000
Aroclor-1262	500	1	6.92	6.82	7.02	563142000
		2	7.18	7.08	7.28	401024000
		3	7.70	7.60	7.80	359470000
		4	7.76	7.66	7.86	666922000
		5	8.26	8.16	8.36	283268000
Aroclor-1268	500	1	7.70	7.60	7.80	1169220000
		2	7.76	7.66	7.86	1148130000
		3	7.97	7.87	8.07	891866000
		4	8.26	8.16	8.36	329470000
		5	8.56	8.46	8.66	2772300000