

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112911.D **Time Analyzed:** 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.751	4.655	4.855	529.280	500.000	5.9
Aroclor-1016-2	4.770	4.674	4.874	529.210	500.000	5.8
Aroclor-1016-3	4.827	4.731	4.931	519.380	500.000	3.9
Aroclor-1016-4	4.947	4.851	5.051	522.620	500.000	4.5
Aroclor-1016-5	5.203	5.108	5.308	552.640	500.000	10.5
Aroclor-1260-1	6.240	6.145	6.345	523.500	500.000	4.7
Aroclor-1260-2	6.430	6.333	6.533	511.640	500.000	2.3
Aroclor-1260-3	6.795	6.700	6.900	536.970	500.000	7.4
Aroclor-1260-4	7.054	6.959	7.159	512.750	500.000	2.6
Aroclor-1260-5	7.298	7.203	7.403	504.960	500.000	1.0
Decachlorobiphenyl	8.687	8.593	8.793	47.260	50.000	-5.5
Tetrachloro-m-xylene	3.666	3.569	3.769	54.550	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112911.D **Time Analyzed:** 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	491.180	500.000	-1.8
Aroclor-1016-2	4.751	4.656	4.856	490.980	500.000	-1.8
Aroclor-1016-3	4.926	4.831	5.031	482.590	500.000	-3.5
Aroclor-1016-4	4.968	4.873	5.073	484.560	500.000	-3.1
Aroclor-1016-5	5.180	5.085	5.285	498.410	500.000	-0.3
Aroclor-1260-1	6.208	6.113	6.313	497.400	500.000	-0.5
Aroclor-1260-2	6.396	6.302	6.502	500.970	500.000	0.2
Aroclor-1260-3	6.547	6.453	6.653	517.790	500.000	3.6
Aroclor-1260-4	7.017	6.923	7.123	515.120	500.000	3.0
Aroclor-1260-5	7.259	7.164	7.364	532.890	500.000	6.6
Decachlorobiphenyl	8.633	8.540	8.740	51.030	50.000	2.1
Tetrachloro-m-xylene	3.659	3.562	3.762	52.400	50.000	4.8