

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074348.D **Time Analyzed:** 15:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.712	5.912	539.370	500.000	7.9
Aroclor-1016-2	5.831	5.733	5.933	519.350	500.000	3.9
Aroclor-1016-3	5.894	5.796	5.996	536.400	500.000	7.3
Aroclor-1016-4	5.991	5.893	6.093	538.480	500.000	7.7
Aroclor-1016-5	6.283	6.186	6.386	527.750	500.000	5.6
Aroclor-1260-1	7.400	7.303	7.503	507.600	500.000	1.5
Aroclor-1260-2	7.653	7.555	7.755	482.260	500.000	-3.5
Aroclor-1260-3	8.011	7.913	8.113	488.950	500.000	-2.2
Aroclor-1260-4	8.239	8.142	8.342	492.340	500.000	-1.5
Aroclor-1260-5	8.565	8.468	8.668	476.780	500.000	-4.6
Decachlorobiphenyl	10.438	10.338	10.538	48.810	50.000	-2.4
Tetrachloro-m-xylene	4.657	4.560	4.760	55.460	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074348.D **Time Analyzed:** 15:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.806	5.006	574.110	500.000	14.8
Aroclor-1016-2	4.959	4.863	5.063	586.940	500.000	17.4
Aroclor-1016-3	5.078	4.983	5.183	583.830	500.000	16.8
Aroclor-1016-4	5.119	5.023	5.223	582.340	500.000	16.5
Aroclor-1016-5	5.334	5.239	5.439	587.300	500.000	17.5
Aroclor-1260-1	6.551	6.455	6.655	541.220	500.000	8.2
Aroclor-1260-2	6.705	6.609	6.809	539.610	500.000	7.9
Aroclor-1260-3	6.915	6.819	7.019	553.120	500.000	10.6
Aroclor-1260-4	7.175	7.079	7.279	545.680	500.000	9.1
Aroclor-1260-5	7.414	7.318	7.518	545.310	500.000	9.1
Decachlorobiphenyl	8.820	8.726	8.926	51.600	50.000	3.2
Tetrachloro-m-xylene	3.800	3.705	3.905	59.690	50.000	19.4