

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074362.D **Time Analyzed:** 20:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.810	5.712	5.912	527.610	500.000	5.5
Aroclor-1016-2	5.832	5.733	5.933	533.020	500.000	6.6
Aroclor-1016-3	5.895	5.796	5.996	544.910	500.000	9.0
Aroclor-1016-4	5.992	5.893	6.093	540.630	500.000	8.1
Aroclor-1016-5	6.285	6.186	6.386	533.180	500.000	6.6
Aroclor-1260-1	7.402	7.303	7.503	509.770	500.000	2.0
Aroclor-1260-2	7.654	7.555	7.755	493.210	500.000	-1.4
Aroclor-1260-3	8.013	7.913	8.113	497.710	500.000	-0.5
Aroclor-1260-4	8.241	8.142	8.342	500.220	500.000	0.0
Aroclor-1260-5	8.567	8.468	8.668	482.520	500.000	-3.5
Decachlorobiphenyl	10.437	10.338	10.538	49.270	50.000	-1.5
Tetrachloro-m-xylene	4.659	4.560	4.760	55.880	50.000	11.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074362.D **Time Analyzed:** 20:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.806	5.006	589.560	500.000	17.9
Aroclor-1016-2	4.961	4.863	5.063	592.870	500.000	18.6
Aroclor-1016-3	5.080	4.983	5.183	589.270	500.000	17.9
Aroclor-1016-4	5.121	5.023	5.223	576.210	500.000	15.2
Aroclor-1016-5	5.334	5.239	5.439	553.350	500.000	10.7
Aroclor-1260-1	6.552	6.455	6.655	560.130	500.000	12.0
Aroclor-1260-2	6.706	6.609	6.809	552.530	500.000	10.5
Aroclor-1260-3	6.916	6.819	7.019	568.070	500.000	13.6
Aroclor-1260-4	7.176	7.079	7.279	559.600	500.000	11.9
Aroclor-1260-5	7.415	7.318	7.518	564.360	500.000	12.9
Decachlorobiphenyl	8.820	8.726	8.926	53.490	50.000	7.0
Tetrachloro-m-xylene	3.801	3.705	3.905	59.430	50.000	18.9