

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031476.D **Time Analyzed:** 09:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.295	9.200	9.400	718.620	712.500	0.9
2,4-D	8.420	8.324	8.524	817.150	705.000	15.9
2,4-DCAA	7.291	7.195	7.395	732.310	750.000	-2.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031476.D **Time Analyzed:** 09:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.896	9.803	10.003	656.000	712.500	-7.9
2,4-D	9.000	8.906	9.106	599.570	705.000	-15.0
2,4-DCAA	7.742	7.649	7.849	655.150	750.000	-12.6