

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031482.D **Time Analyzed:** 12:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.292	9.200	9.400	697.840	712.500	-2.1
2,4-D	8.417	8.324	8.524	673.700	705.000	-4.4
2,4-DCAA	7.289	7.195	7.395	714.250	750.000	-4.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031482.D **Time Analyzed:** 12:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.896	9.803	10.003	661.360	712.500	-7.2
2,4-D	9.002	8.906	9.106	622.860	705.000	-11.7
2,4-DCAA	7.744	7.649	7.849	656.580	750.000	-12.5