

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089914.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.069	8.971	9.171	41.490	50.000	-17.0
Endrin	6.572	6.472	6.672	42.180	50.000	-15.6
gamma-BHC (Lindane)	4.329	4.228	4.428	46.180	50.000	-7.6
Heptachlor	4.927	4.827	5.027	43.520	50.000	-13.0
Heptachlor epoxide	5.689	5.589	5.789	44.170	50.000	-11.7
Methoxychlor	7.491	7.392	7.592	41.550	50.000	-16.9
Tetrachloro-m-xylene	3.549	3.448	3.648	45.100	50.000	-9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089914.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.067	7.968	8.168	48.370	50.000	-3.3
Endrin	5.786	5.685	5.885	57.310	50.000	14.6
gamma-BHC (Lindane)	3.727	3.626	3.826	54.950	50.000	9.9
Heptachlor	4.080	3.979	4.179	52.420	50.000	4.8
Heptachlor epoxide	4.869	4.769	4.969	53.730	50.000	7.5
Methoxychlor	6.750	6.650	6.850	52.950	50.000	5.9
Tetrachloro-m-xylene	2.879	2.778	2.978	54.350	50.000	8.7