

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089935.D **Time Analyzed:** 20:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.068	8.971	9.171	44.170	50.000	-11.7
Endrin	6.571	6.472	6.672	44.210	50.000	-11.6
gamma-BHC (Lindane)	4.328	4.228	4.428	46.500	50.000	-7.0
Heptachlor	4.927	4.827	5.027	45.060	50.000	-9.9
Heptachlor epoxide	5.688	5.589	5.789	45.890	50.000	-8.2
Methoxychlor	7.490	7.392	7.592	43.220	50.000	-13.6
Tetrachloro-m-xylene	3.548	3.448	3.648	45.570	50.000	-8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089935.D **Time Analyzed:** 20:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.065	7.968	8.168	49.970	50.000	-0.1
Endrin	5.785	5.685	5.885	58.980	50.000	18.0
gamma-BHC (Lindane)	3.727	3.626	3.826	55.740	50.000	11.5
Heptachlor	4.080	3.979	4.179	54.350	50.000	8.7
Heptachlor epoxide	4.869	4.769	4.969	54.750	50.000	9.5
Methoxychlor	6.748	6.650	6.850	57.120	50.000	14.2
Tetrachloro-m-xylene	2.878	2.778	2.978	55.190	50.000	10.4