

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/16/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089947.D **Time Analyzed:** 00:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.067	8.971	9.171	43.440	50.000	-13.1
Endrin	6.572	6.472	6.672	45.050	50.000	-9.9
gamma-BHC (Lindane)	4.329	4.228	4.428	47.250	50.000	-5.5
Heptachlor	4.927	4.827	5.027	45.500	50.000	-9.0
Heptachlor epoxide	5.688	5.589	5.789	45.580	50.000	-8.8
Methoxychlor	7.490	7.392	7.592	43.920	50.000	-12.2
Tetrachloro-m-xylene	3.549	3.448	3.648	46.330	50.000	-7.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2836
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/31/2025 07/31/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/16/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089947.D **Time Analyzed:** 00:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.066	7.968	8.168	49.230	50.000	-1.5
Endrin	5.786	5.685	5.885	63.120	50.000	26.2
gamma-BHC (Lindane)	3.727	3.626	3.826	56.770	50.000	13.5
Heptachlor	4.079	3.979	4.179	55.190	50.000	10.4
Heptachlor epoxide	4.869	4.769	4.969	55.650	50.000	11.3
Methoxychlor	6.749	6.650	6.850	57.690	50.000	15.4
Tetrachloro-m-xylene	2.879	2.778	2.978	55.970	50.000	11.9