

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2836</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ENTACT</u>	DataFile :	<u>PD089920.D</u>

	Parameter	Spike	Sample			Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result	Units							High	RPD
Lab Sample ID:	Q2832-02MS	Client Sample ID:		TG-S01MS								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	5.10	ug/L	102				30 (60)	150 (152)	
	Heptachlor	5	0	4.60	ug/L	92				30 (56)	150 (147)	
	Heptachlor epoxide	5	0	4.70	ug/L	94				30 (77)	150 (143)	
	Endrin	5	0	4.70	ug/L	94				30 (76)	150 (144)	
	Methoxychlor	5	0	4.30	ug/L	86				30 (70)	150 (142)	
Lab Sample ID:	Q2832-02MS	Client Sample ID:		TG-S01MS								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	5.90	ug/L	118				30 (60)	150 (152)	
	Heptachlor	5	0	5.40	ug/L	108				30 (56)	150 (147)	
	Heptachlor epoxide	5	0	5.80	ug/L	116				30 (77)	150 (143)	
	Endrin	5	0	5.60	ug/L	112				30 (76)	150 (144)	
	Methoxychlor	5	0	5.30	ug/L	106				30 (70)	150 (142)	