

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |               |                           |                   |
|-----------------|---------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q2836</u>  | <b>Analytical Method:</b> | <u>8081B</u>      |
| <b>Client:</b>  | <u>ENTACT</u> | <b>Datafile :</b>         | <u>PD089937.D</u> |

| Lab Sample ID            | Parameter           | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|--------------------------|---------------------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|                          |                     |       |        |       |     |     |      | Qual | Low     | High      |     |
| PB169263BS<br>(Column 1) | gamma-BHC (Lindane) | 0.5   | 0.48   | ug/L  | 96  |     |      |      | 40 (82) | 140 (129) |     |
|                          | Heptachlor          | 0.5   | 0.47   | ug/L  | 94  |     |      |      | 40 (79) | 140 (127) |     |
|                          | Heptachlor epoxide  | 0.5   | 0.47   | ug/L  | 94  |     |      |      | 40 (81) | 140 (124) |     |
|                          | Endrin              | 0.5   | 0.46   | ug/L  | 91  |     |      |      | 40 (81) | 140 (128) |     |
|                          | Methoxychlor        | 0.5   | 0.44   | ug/L  | 87  |     |      |      | 40 (78) | 140 (108) |     |
| PB169263BS<br>(Column 2) | gamma-BHC (Lindane) | 0.5   | 0.58   | ug/L  | 115 |     |      |      | 40 (82) | 140 (129) |     |
|                          | Heptachlor          | 0.5   | 0.56   | ug/L  | 113 |     |      |      | 40 (79) | 140 (127) |     |
|                          | Heptachlor epoxide  | 0.5   | 0.57   | ug/L  | 115 |     |      |      | 40 (81) | 140 (124) |     |
|                          | Endrin              | 0.5   | 0.57   | ug/L  | 114 |     |      |      | 40 (81) | 140 (128) |     |
|                          | Methoxychlor        | 0.5   | 0.56   | ug/L  | 113 |     |      |      | 40 (78) | 140 (108) |     |