

Surrogate Summary

SDG No.: Q2836

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2836-01	WC-A2-15-G	1,2-Dichloroethane-d4	50	56.8	114		70 (74)	130 (125)
		Dibromofluoromethane	50	42.0	84		70 (75)	130 (124)
		Toluene-d8	50	46.0	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.1	88		70 (77)	130 (121)
Q2836-05	WC-A2-16-G	1,2-Dichloroethane-d4	50	53.8	108		70 (74)	130 (125)
		Dibromofluoromethane	50	41.0	82		70 (75)	130 (124)
		Toluene-d8	50	45.0	90		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.6	87		70 (77)	130 (121)
Q2836-09	WC-A2-17-G	1,2-Dichloroethane-d4	50	54.4	109		70 (74)	130 (125)
		Dibromofluoromethane	50	41.5	83		70 (75)	130 (124)
		Toluene-d8	50	45.3	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.4	87		70 (77)	130 (121)
Q2836-13	WC-A5-02-G	1,2-Dichloroethane-d4	50	57.5	115		70 (74)	130 (125)
		Dibromofluoromethane	50	41.1	82		70 (75)	130 (124)
		Toluene-d8	50	46.5	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.4	91		70 (77)	130 (121)

Surrogate Summary

SDG No.: Q2836

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VN0814WBL01	VN0814WBL01	1,2-Dichloroethane-d4	50	52.9	106		70 (74)	130 (125)
		Dibromofluoromethane	50	44.2	88		70 (75)	130 (124)
		Toluene-d8	50	44.2	88		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.7	87		70 (77)	130 (121)
VN0814WBS01	VN0814WBS01	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	50.7	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106		70 (77)	130 (121)
VN0814WBSD0	VN0814WBSD01	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99		70 (75)	130 (124)
		Toluene-d8	50	49.7	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.6	99		70 (77)	130 (121)

