



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q2837

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.92	6.0	99	65 - 135	P	08/14/2025	12:23	LB136837
	Copper	21.4	20.0	107	65 - 135	P	08/14/2025	12:23	LB136837
	Lead	12.4	12.0	103	65 - 135	P	08/14/2025	12:23	LB136837
	Nickel	40.2	40.0	100	65 - 135	P	08/14/2025	12:23	LB136837
	Zinc	43.0	40.0	108	65 - 135	P	08/14/2025	12:23	LB136837
CRA	Mercury	0.19	0.2	94	70 - 130	CV	08/15/2025	12:16	LB136836