

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT

SDG No.: Q2837

Contract: ENTA05

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	2.91	1.0	291	-5	7	08/14/2025	12:27	LB136837
	Copper	7.21	2.0	360	-18	22	08/14/2025	12:27	LB136837
	Lead	-5.21			-12	12	08/14/2025	12:27	LB136837
	Nickel	5.40	2.0	270	-38	42	08/14/2025	12:27	LB136837
	Zinc	3.33			-40	40	08/14/2025	12:27	LB136837
ICSAB01	Cadmium	989	972	102	826	1120	08/14/2025	12:38	LB136837
	Copper	490	511	96	434	588	08/14/2025	12:38	LB136837
	Lead	42.6	49.0	87	37	61	08/14/2025	12:38	LB136837
	Nickel	987	954	104	810	1100	08/14/2025	12:38	LB136837
	Zinc	1040	952	109	809	1095	08/14/2025	12:38	LB136837