

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2837

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169232BS							
Cadmium	ug/L	100	97.0		97	88 - 113	P
Copper	ug/L	150	158		105	86 - 114	P
Lead	ug/L	500	488		98	86 - 113	P
Nickel	ug/L	250	251		100	88 - 113	P
Zinc	ug/L	100	106		106	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2837

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169249BS Mercury	ug/L	4.0	4.19		105	82 - 119	CV