

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2837
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074406.D **Time Analyzed:** 09:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.712	5.912	479.870	500.000	-4.0
Aroclor-1016-2	5.827	5.733	5.933	502.730	500.000	0.5
Aroclor-1016-3	5.890	5.796	5.996	500.440	500.000	0.1
Aroclor-1016-4	5.987	5.893	6.093	503.910	500.000	0.8
Aroclor-1016-5	6.280	6.186	6.386	500.100	500.000	0.0
Aroclor-1260-1	7.398	7.303	7.503	474.450	500.000	-5.1
Aroclor-1260-2	7.650	7.555	7.755	455.030	500.000	-9.0
Aroclor-1260-3	8.008	7.913	8.113	463.430	500.000	-7.3
Aroclor-1260-4	8.236	8.142	8.342	466.790	500.000	-6.6
Aroclor-1260-5	8.562	8.468	8.668	447.650	500.000	-10.5
Decachlorobiphenyl	10.433	10.338	10.538	46.020	50.000	-8.0
Tetrachloro-m-xylene	4.654	4.560	4.760	51.940	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2837
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074406.D **Time Analyzed:** 09:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.897	4.806	5.006	556.370	500.000	11.3
Aroclor-1016-2	4.956	4.863	5.063	570.700	500.000	14.1
Aroclor-1016-3	5.076	4.983	5.183	556.650	500.000	11.3
Aroclor-1016-4	5.116	5.023	5.223	571.310	500.000	14.3
Aroclor-1016-5	5.330	5.239	5.439	586.740	500.000	17.3
Aroclor-1260-1	6.547	6.455	6.655	552.380	500.000	10.5
Aroclor-1260-2	6.701	6.609	6.809	500.850	500.000	0.2
Aroclor-1260-3	6.911	6.819	7.019	545.710	500.000	9.1
Aroclor-1260-4	7.171	7.079	7.279	533.150	500.000	6.6
Aroclor-1260-5	7.410	7.318	7.518	529.380	500.000	5.9
Decachlorobiphenyl	8.815	8.726	8.926	50.710	50.000	1.4
Tetrachloro-m-xylene	3.798	3.705	3.905	57.730	50.000	15.5