

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2837
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074420.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.712	5.912	509.250	500.000	1.9
Aroclor-1016-2	5.833	5.733	5.933	503.490	500.000	0.7
Aroclor-1016-3	5.895	5.796	5.996	512.030	500.000	2.4
Aroclor-1016-4	5.993	5.893	6.093	512.460	500.000	2.5
Aroclor-1016-5	6.285	6.186	6.386	508.660	500.000	1.7
Aroclor-1260-1	7.402	7.303	7.503	489.160	500.000	-2.2
Aroclor-1260-2	7.655	7.555	7.755	470.920	500.000	-5.8
Aroclor-1260-3	8.013	7.913	8.113	478.770	500.000	-4.2
Aroclor-1260-4	8.241	8.142	8.342	482.050	500.000	-3.6
Aroclor-1260-5	8.567	8.468	8.668	459.280	500.000	-8.1
Decachlorobiphenyl	10.440	10.338	10.538	47.210	50.000	-5.6
Tetrachloro-m-xylene	4.659	4.560	4.760	52.870	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2837
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074420.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.806	5.006	571.890	500.000	14.4
Aroclor-1016-2	4.960	4.863	5.063	592.020	500.000	18.4
Aroclor-1016-3	5.081	4.983	5.183	577.180	500.000	15.4
Aroclor-1016-4	5.122	5.023	5.223	565.300	500.000	13.1
Aroclor-1016-5	5.335	5.239	5.439	622.420	500.000	24.5
Aroclor-1260-1	6.552	6.455	6.655	541.450	500.000	8.3
Aroclor-1260-2	6.706	6.609	6.809	537.610	500.000	7.5
Aroclor-1260-3	6.915	6.819	7.019	577.120	500.000	15.4
Aroclor-1260-4	7.176	7.079	7.279	551.230	500.000	10.2
Aroclor-1260-5	7.415	7.318	7.518	549.180	500.000	9.8
Decachlorobiphenyl	8.820	8.726	8.926	51.020	50.000	2.0
Tetrachloro-m-xylene	3.803	3.705	3.905	59.140	50.000	18.3