

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP074166.D	01 Aug 2025 11:33	YP\AJ	Ok
2	I.BLK	PP074167.D	01 Aug 2025 11:49	YP\AJ	Ok
3	AR1660ICC1000	PP074168.D	01 Aug 2025 12:05	YP\AJ	Ok
4	AR1660ICC750	PP074169.D	01 Aug 2025 12:22	YP\AJ	Ok
5	AR1660ICC500	PP074170.D	01 Aug 2025 12:38	YP\AJ	Ok
6	AR1660ICC250	PP074171.D	01 Aug 2025 12:54	YP\AJ	Ok,M
7	AR1660ICC050	PP074172.D	01 Aug 2025 13:42	YP\AJ	Ok,M
8	AR1221ICC500	PP074173.D	01 Aug 2025 13:58	YP\AJ	Ok
9	AR1232ICC500	PP074174.D	01 Aug 2025 14:15	YP\AJ	Ok,M
10	AR1242ICC1000	PP074175.D	01 Aug 2025 14:31	YP\AJ	Ok,M
11	AR1242ICC750	PP074176.D	01 Aug 2025 14:47	YP\AJ	Ok,M
12	AR1242ICC500	PP074177.D	01 Aug 2025 15:03	YP\AJ	Ok,M
13	AR1242ICC250	PP074178.D	01 Aug 2025 15:19	YP\AJ	Ok,M
14	AR1242ICC050	PP074179.D	01 Aug 2025 15:36	YP\AJ	Ok,M
15	AR1248ICC1000	PP074180.D	01 Aug 2025 16:24	YP\AJ	Not Ok
16	AR1248ICC750	PP074181.D	01 Aug 2025 16:41	YP\AJ	Not Ok
17	AR1248ICC500	PP074182.D	01 Aug 2025 16:57	YP\AJ	Ok,M
18	AR1248ICC250	PP074183.D	01 Aug 2025 17:13	YP\AJ	Not Ok
19	AR1248ICC050	PP074184.D	01 Aug 2025 17:45	YP\AJ	Not Ok
20	AR1254ICC1000	PP074185.D	01 Aug 2025 18:02	YP\AJ	Ok,M
21	AR1254ICC750	PP074186.D	01 Aug 2025 18:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP074187.D	01 Aug 2025 18:34	YP\AJ	Ok,M
23	AR1254ICC250	PP074188.D	01 Aug 2025 18:50	YP\AJ	Ok,M
24	AR1254ICC050	PP074189.D	01 Aug 2025 19:23	YP\AJ	Ok,M
25	AR1262ICC500	PP074190.D	01 Aug 2025 19:39	YP\AJ	Ok
26	AR1268ICC1000	PP074191.D	01 Aug 2025 19:55	YP\AJ	Not Ok
27	AR1268ICC750	PP074192.D	01 Aug 2025 20:11	YP\AJ	Not Ok
28	AR1268ICC500	PP074193.D	01 Aug 2025 20:28	YP\AJ	Ok
29	AR1268ICC250	PP074194.D	01 Aug 2025 20:44	YP\AJ	Not Ok
30	AR1268ICC050	PP074195.D	01 Aug 2025 21:00	YP\AJ	Not Ok
31	PP080125ICV500	PP074196.D	01 Aug 2025 21:16	YP\AJ	Ok
32	AR1242ICV500	PP074197.D	01 Aug 2025 22:05	YP\AJ	Ok,M
33	AR1248ICV500	PP074198.D	01 Aug 2025 22:37	YP\AJ	Not Ok
34	AR1254ICV500	PP074199.D	01 Aug 2025 23:10	YP\AJ	Ok,M
35	AR1268ICV500	PP074200.D	01 Aug 2025 23:42	YP\AJ	Not Ok
36	DDT ANALOG	PP074201.D	02 Aug 2025 00:15	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081525

Review By	yogesh	Review On	8/15/2025 1:04:15 PM
Supervise By	mohammad	Supervise On	8/19/2025 1:26:40 AM
SubDirectory	PP081525	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP074405.D	15 Aug 2025 08:47	YP\AJ	Ok
2	AR1660CCC500	PP074406.D	15 Aug 2025 09:03	YP\AJ	Ok
3	AR1242CCC500	PP074407.D	15 Aug 2025 09:19	YP\AJ	Ok,M
4	AR1254CCC500	PP074408.D	15 Aug 2025 09:35	YP\AJ	Ok,M
5	I.BLK	PP074409.D	15 Aug 2025 09:52	YP\AJ	Ok
6	DDT ANALOG	PP074410.D	15 Aug 2025 10:08	YP\AJ	Ok
7	PB169241BL	PP074411.D	15 Aug 2025 10:24	YP\AJ	Ok
8	PB169241BS	PP074412.D	15 Aug 2025 10:40	YP\AJ	Ok,M
9	PB169241BSD	PP074413.D	15 Aug 2025 10:57	YP\AJ	Ok,M
10	Q2837-01	PP074414.D	15 Aug 2025 11:13	YP\AJ	Ok,M
11	Q2840-03	PP074415.D	15 Aug 2025 11:29	YP\AJ	Not Ok
12	Q2849-01	PP074416.D	15 Aug 2025 11:45	YP\AJ	Ok,M
13	Q2865-02	PP074417.D	15 Aug 2025 12:28	YP\AJ	Ok,M
14	PB169258BL	PP074418.D	15 Aug 2025 13:01	YP\AJ	Ok
15	PB169258BS	PP074419.D	15 Aug 2025 13:17	YP\AJ	Ok
16	AR1660CCC500	PP074420.D	15 Aug 2025 14:22	YP\AJ	Ok
17	AR1242CCC500	PP074421.D	15 Aug 2025 14:38	YP\AJ	Ok,M
18	AR1254CCC500	PP074422.D	15 Aug 2025 14:55	YP\AJ	Ok,M
19	I.BLK	PP074423.D	15 Aug 2025 15:11	YP\AJ	Ok
20	PB169260BL	PP074424.D	15 Aug 2025 15:27	YP\AJ	Ok
21	PB169260BS	PP074425.D	15 Aug 2025 15:43	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081525

Review By	yogesh	Review On	8/15/2025 1:04:15 PM
Supervise By	mohammad	Supervise On	8/19/2025 1:26:40 AM
SubDirectory	PP081525	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2871-01	PP074426.D	15 Aug 2025 16:00	YP\AJ	Ok
23	Q2871-01MS	PP074427.D	15 Aug 2025 16:16	YP\AJ	Ok
24	Q2871-01MSD	PP074428.D	15 Aug 2025 16:32	YP\AJ	Ok
25	Q2871-04	PP074429.D	15 Aug 2025 16:48	YP\AJ	Ok
26	Q2871-07	PP074430.D	15 Aug 2025 17:05	YP\AJ	Ok
27	Q2872-01	PP074431.D	15 Aug 2025 17:21	YP\AJ	Ok
28	Q2873-01	PP074432.D	15 Aug 2025 17:37	YP\AJ	Ok
29	Q2873-04	PP074433.D	15 Aug 2025 17:53	YP\AJ	Ok
30	AR1660CCC500	PP074434.D	15 Aug 2025 18:58	YP\AJ	Ok
31	AR1242CCC500	PP074435.D	15 Aug 2025 19:47	YP\AJ	Ok,M
32	AR1254CCC500	PP074436.D	15 Aug 2025 20:03	YP\AJ	Ok,M
33	I.BLK	PP074437.D	15 Aug 2025 20:20	YP\AJ	Ok

M : Manual Integration