

Surrogate Summary

SW-846

SDG No.: Q2837

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169230BL	PB169230BL	2-Fluorophenol	150	124	83		15 (23)	110 (138)
		Phenol-d6	150	118	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.2	77		30 (67)	130 (132)
		2-Fluorobiphenyl	100	77.3	77		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	128	85		15 (44)	110 (137)
		Terphenyl-d14	100	74.2	74		30 (42)	130 (152)
PB169230BS	PB169230BS	2-Fluorophenol	150	113	75		15 (23)	110 (138)
		Phenol-d6	150	110	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	69.6	70		30 (67)	130 (132)
		2-Fluorobiphenyl	100	68.6	69		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	111	74		15 (44)	110 (137)
		Terphenyl-d14	100	65.8	66		30 (42)	130 (152)
PB169230BSD	PB169230BSD	2-Fluorophenol	150	124	83		15 (23)	110 (138)
		Phenol-d6	150	123	82		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.4	77		30 (67)	130 (132)
		2-Fluorobiphenyl	100	76.8	77		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	125	83		15 (44)	110 (137)
		Terphenyl-d14	100	77.3	77		30 (42)	130 (152)
Q2837-01	TW-WTS-13	2-Fluorophenol	150	65.8	44		15 (23)	110 (138)
		Phenol-d6	150	40.5	27		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.4	93		30 (67)	130 (132)
		2-Fluorobiphenyl	100	91.3	91		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	127	85		15 (44)	110 (137)
		Terphenyl-d14	100	69.9	70		30 (42)	130 (152)