

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2837

Analytical Method: 8270E

Client: ENTACT

DataFile: BP025425.D

| Lab Sample ID | Parameter              | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits  |           |         |
|---------------|------------------------|-------|--------|------|-----|-----|------|------|---------|-----------|---------|
|               |                        |       |        |      |     |     |      | Qual | Low     | High      | RPD     |
| PB169230BSD   | Phenol                 | 50    | 44.9   | ug/L | 90  | 11  |      |      | 20 (66) | 160 (118) | 20 (20) |
|               | 1,4-Dichlorobenzene    | 50    | 41.7   | ug/L | 83  | 10  |      |      | 70 (76) | 130 (103) | 20 (20) |
|               | 1,2,4-Trichlorobenzene | 50    | 42.7   | ug/L | 85  | 10  |      |      | 70 (41) | 130 (115) | 20 (20) |
|               | Naphthalene            | 50    | 43.1   | ug/L | 86  | 11  |      |      | 70 (64) | 130 (107) | 20 (20) |