

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2837

Client ID : SSTDCCC040

Date Analyzed: 08/15/2025

Lab File ID: BP025422.D

Time Analyzed: 10:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	163210	7.79	699113	10.55	458301	14.41
UPPER LIMIT	326420	8.29	1398230	11.054	916602	14.907
LOWER LIMIT	81605	7.29	349557	10.054	229151	13.907
EPA SAMPLE NO.						
01 PB169230BL	154254	7.79	595195	10.56	376083	14.40
02 PB169230BS	169971	7.79	670319	10.56	424018	14.40
03 PB169230BSD	168072	7.78	666741	10.55	424927	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.