

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: BP025421.D  
Instrument ID: BNA\_P

Contract: ENTA05  
SDG NO.: Q2837  
DFTPP Injection Date: 08/15/2025  
DFTPP Injection Time: 09:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 ( 1.8 ) 1
69	Mass 69 relative abundance	28.2
70	Less than 2.0% of mass 69	0.1 ( 0.5 ) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 ( 19.5 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025422.D	08/15/2025	10:36
PB169230BL	PB169230BL	BP025423.D	08/15/2025	11:17
PB169230BS	PB169230BS	BP025424.D	08/15/2025	11:59
PB169230BSD	PB169230BSD	BP025425.D	08/15/2025	13:26