

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2837</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/19/2025 10:44</u>
Lab File ID:	<u>BP025470.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.204	1.229		2.1	
Phenol-d6	1.544	1.586		2.7	
Phenol	1.620	1.652		2.0	20.0
1,4-Dichlorobenzene	1.532	1.544		0.8	20.0
Nitrobenzene-d5	0.395	0.403		2.0	
1,2,4-Trichlorobenzene	0.340	0.335		-1.5	
Naphthalene	1.040	1.035		-0.5	
2-Fluorobiphenyl	1.463	1.448		-1.0	
2,4,6-Tribromophenol	0.269	0.273		1.5	
Terphenyl-d14	1.093	1.093		0.0	

All other compounds must meet a minimum RRF of 0.010.