

Surrogate Summary

SW-846

SDG No.: Q2837

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169244BL	PB169244BL	2-Fluorophenol	150	129	86		15 (23)	110 (138)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.8	78		30 (67)	130 (132)
		2-Fluorobiphenyl	100	75.3	75		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	88		15 (44)	110 (137)
		Terphenyl-d14	100	76.5	76		30 (42)	130 (152)
PB169244BS	PB169244BS	2-Fluorophenol	150	138	92		15 (23)	110 (138)
		Phenol-d6	150	136	90		15 (10)	110 (134)
		Nitrobenzene-d5	100	82.4	82		30 (67)	130 (132)
		2-Fluorobiphenyl	100	81.1	81		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	138	92		15 (44)	110 (137)
		Terphenyl-d14	100	81.9	82		30 (42)	130 (152)
PB169244BSD	PB169244BSD	2-Fluorophenol	150	141	94		15 (23)	110 (138)
		Phenol-d6	150	141	94		15 (10)	110 (134)
		Nitrobenzene-d5	100	84.6	85		30 (67)	130 (132)
		2-Fluorobiphenyl	100	79.2	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	144	96		15 (44)	110 (137)
		Terphenyl-d14	100	84.2	84		30 (42)	130 (152)
Q2837-01	TW-WTS-13	2-Fluorophenol	150	65.8	44		15 (23)	110 (138)
		Phenol-d6	150	40.5	27		15 (10)	110 (134)
		Nitrobenzene-d5	100	93.4	93		30 (67)	130 (132)
		2-Fluorobiphenyl	100	91.3	91		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	127	85		15 (44)	110 (137)
		Terphenyl-d14	100	69.9	70		30 (42)	130 (152)