

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2837

Client ID : SSTDCCC040

Date Analyzed: 08/19/2025

Lab File ID: BP025470.D

Time Analyzed: 10:44

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	152340	7.784	634765	10.56	424501	14.40
UPPER LIMIT	304680	8.284	1269530	11.06	849002	14.895
LOWER LIMIT	76170	7.284	317383	10.06	212251	13.895
EPA SAMPLE NO.						
01 PB169244BL	171241	7.78	669280	10.56	431883	14.41
02 PB169244BS	161110	7.79	644957	10.56	415960	14.40
03 PB169244BSD	172374	7.79	704478	10.56	488785	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.