

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2837

Client ID: SSTDCCC040

Date Analyzed: 08/18/2025

Lab File ID: BF143429.D

Time Analyzed: 09:06

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	444924	11.445	223862	14.086	265880	15.574
UPPER LIMIT	889848	11.945	447724	14.586	531760	16.074
LOWER LIMIT	222462	10.945	111931	13.586	132940	15.074
EPA SAMPLE NO.						
TW-WTS-13	263021	11.45	240460	14.08	295102	15.57

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.