

## Utsav Patel

---

**From:** Yazmeen Gomez  
**Sent:** Thursday, August 14, 2025 10:30 AM  
**To:** Ynfante, John; Sohil Jodhani  
**Cc:** Mohammad Ahmed  
**Subject:** RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.  
**Attachments:** quant2841.pdf; quant2842.pdf

John,

The lab informed me –

Q2841- 42 sample screened with 8260 method as they are usually highly contaminated samples for Trichloroethene

Q2841-01 (MW-06-6.5-08122025) at 200x

Q2842-07(MW-17B-55.5-081225-SIM) 200x





Sample can be analyzed with 8260 under low dilution

Sim analysis is not possible as it may contaminate instrument

Please see attached.

**Best Regards,**



**Yazmeen** Gomez  
**Sr. Project Manager**  
**An Alliance Technical Group Company**  
**Main:** 908-789-8900  
**Direct:** 908-728-3147  
**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092  
[www.alliancetg.com](http://www.alliancetg.com)    

---

**From:** Ynfante, John <John.Ynfante@jacobs.com>  
**Sent:** Wednesday, August 13, 2025 3:23 PM  
**To:** Sohil Jodhani <Sohil.Jodhani@alliancetg.com>  
**Cc:** Yazmeen Gomez <Yazmeen.Gomez@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>  
**Subject:** RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Thanks Sohil

**From:** Sohil Jodhani <[Sohil.Jodhani@alliancetg.com](mailto:Sohil.Jodhani@alliancetg.com)>  
**Sent:** Wednesday, August 13, 2025 1:23 PM  
**To:** Ynfante, John <[John.Ynfante@jacobs.com](mailto:John.Ynfante@jacobs.com)>  
**Cc:** Yazmeen Gomez <[Yazmeen.Gomez@AllianceTG.com](mailto:Yazmeen.Gomez@AllianceTG.com)>; Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>;  
 Nimisha Pandya <[Nimisha.Pandya@AllianceTG.com](mailto:Nimisha.Pandya@AllianceTG.com)>  
**Subject:** RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,

Sample IDs are updated as per below request. Lab will notify you if in case VOC-SIM (for Vinyl chloride) is not required for the samples based on low level VOA analysis.

Thanks & Regards,



There's a better way.

**Sohil Jodhani** (he/him/his)  
 QA/QC Director  
 Alliance Technical Group, LLC-Newark  
 Main: 908-789-8900  
 Direct: 908-728-3152  
 Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092  
[www.alliancetg.com](http://www.alliancetg.com)    

**From:** Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>  
**Sent:** Wednesday, August 13, 2025 2:10 PM  
**To:** Ynfante, John <[John.Ynfante@jacobs.com](mailto:John.Ynfante@jacobs.com)>; Data-EWR <[Data-EWR@alliancetg.com](mailto:Data-EWR@alliancetg.com)>; Yazmeen Gomez <[yazmeen.gomez@alliancetg.com](mailto:yazmeen.gomez@alliancetg.com)>  
**Cc:** Sohil Jodhani <[Sohil.Jodhani@alliancetg.com](mailto:Sohil.Jodhani@alliancetg.com)>  
**Subject:** Re: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,

yaz is out of the office today, but i will coordinate with lab to make sure we communicate with you and run these analysis accordingly



There's a better way.

**Mohammad Ahmed**  
 Laboratory Director  
 An Alliance Technical Group Company  
 Main: 908-789-8900  
 Direct: 908-728-3151  
 Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092  
[www.alliancetg.com](http://www.alliancetg.com)

**From:** Ynfante, John <[John.Ynfante@jacobs.com](mailto:John.Ynfante@jacobs.com)>

**Sent:** Wednesday, August 13, 2025 2:01 PM

**To:** Data-EWR <[Data-EWR@alliancetg.com](mailto:Data-EWR@alliancetg.com)>; Yazmeen Gomez <[yazmeen.gomez@alliancetg.com](mailto:yazmeen.gomez@alliancetg.com)>

**Subject:** RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Yazmeen,

A couple of comments on this Princeton login for Q2842:

1. I assume your SFAM-SIM method for VOCs can still only report vinyl chloride out of our project list of VOCs – if that is the case then note that sample Q2842-07 (MW-17B-55.5-081225-SIM) and any other SFAM-SIM samples only need to be analyzed if their corresponding 8260D-Low analysis shows the vinyl chloride to be non-detect and of course if the lab is able to analyze them by SIM. I know up to this point the lab has said the concentrations have been too high in the SIM samples to even run the SFAM-SIM on so let me know how it looks on those after your analysts run the regular VOCs run on them.
2. The date part of the sample IDs should have used a MMDDYY format, not MMDDYYYY as is listed on the chain. Please change the “2025” bit to just “25” on all these sample IDs. For example, “RMW-03B-90.4-08122025” should be changed to “RMW-03B-90.4-081225”, etc.

Thanks

- John Y.

---


**From:** [Data-EWR@alliancetg.com](mailto:Data-EWR@alliancetg.com) <[Data-EWR@alliancetg.com](mailto:Data-EWR@alliancetg.com)>

**Sent:** Wednesday, August 13, 2025 9:38 AM

**To:** Ongjoco, Alec <[Alec.Ongjoco@jacobs.com](mailto:Alec.Ongjoco@jacobs.com)>; Dillon, Alexa <[Alexa.Dillon@jacobs.com](mailto:Alexa.Dillon@jacobs.com)>; Lader, Chelsea <[Chelsea.Lader@jacobs.com](mailto:Chelsea.Lader@jacobs.com)>; Holmes, Daniel <[Daniel.Holmes@jacobs.com](mailto:Daniel.Holmes@jacobs.com)>; Reamer, David <[David.Reamer@jacobs.com](mailto:David.Reamer@jacobs.com)>; Ynfante, John <[John.Ynfante@jacobs.com](mailto:John.Ynfante@jacobs.com)>; Murphy, Mary <[Mary.Murphy@jacobs.com](mailto:Mary.Murphy@jacobs.com)>; Warren, Melissa <[Melissa.Warren@jacobs.com](mailto:Melissa.Warren@jacobs.com)>; Asher, Sarah <[Sarah.Asher@jacobs.com](mailto:Sarah.Asher@jacobs.com)>

**Cc:** [yazmeen.gomez@alliancetg.com](mailto:yazmeen.gomez@alliancetg.com)

**Subject:** [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.



To John Ynfante;

Please see the attached Login Summary for the following project, or download the file using your login credentials from the link below.

**Order ID** : Q2842  
**Project ID** : Former Schlumberger STC PTC Site D3868221  
**Download File** : <https://chemtech.net/secureLogin.aspx>  
**Order Date** : 8/13/2025 8:34:00 AM

**Alliance's Project Manager** : YAZMEEN GOMEZ , [yazmeen.gomez@alliancetg.com](mailto:yazmeen.gomez@alliancetg.com) , 908-728-3147  
**Alliance's Sales Executive** : Jordan Hedvat , [jordan.hedvat@alliancetg.com](mailto:jordan.hedvat@alliancetg.com) , 908-728-3144

Thank you for the opportunity to provide you with our services. For any questions please feel free to contact your project manager.

Click Here for our short online customer Survey <http://chemtech.net/ClientSurvey.aspx>.

**Thank you,**

**Alliance Technical Group LLC.**

Notice: The information transmitted in this e-mail message and in any attachments is intended Solely for the attention of the named addressee(s) and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon, this information by persons or entities other than the intended recipient is strictly prohibited and may be unlawful. If you have received this transmission in error, please notify us immediately by return e-mail, and permanently delete this transmission, including attachments if any, from any computer.

■

---

NOTICE - This communication may contain confidential and privileged information that is for the sole use of the intended recipient. Any viewing, copying or distribution of, or reliance on this message by unintended recipients is strictly prohibited. If you have received this message in error, please notify us immediately by replying to the message and deleting it from your computer.

---

NOTICE - This communication may contain confidential and privileged information that is for the sole use of the intended recipient. Any viewing, copying or distribution of, or reliance on this message by unintended recipients is strictly prohibited. If you have received this message in error, please notify us immediately by replying to the message and deleting it from your computer.