

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>JAC005</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q2841</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>08/15/2025</u> <u>08/15/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:40</u> <u>12:30</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF005 = VX047349.D		RRF020 = VX047350.D		RRF050 = VX047351.D		
		RRF100 = VX047352.D		RRF150 = VX047353.D		RRF001 = VX047356.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
Vinyl Chloride	0.651	0.751	0.800	0.729	0.757	0.630	0.720	9.2
1,1-Dichloroethene	0.575	0.664	0.697	0.635	0.651	0.555	0.630	8.6
1,1-Dichloroethane	1.180	1.294	1.356	1.210	1.238	1.052	1.222	8.5
cis-1,2-Dichloroethene	0.758	0.826	0.846	0.769	0.779	0.692	0.778	7
1,1,1-Trichloroethane	1.043	1.151	1.195	1.081	1.106	0.863	1.073	10.8
Benzene	1.502	1.665	1.701	1.516	1.504	1.312	1.533	9.1
1,2-Dichloroethane	0.521	0.594	0.600	0.532	0.528	0.484	0.543	8.3
Trichloroethene	0.383	0.423	0.432	0.390	0.391	0.336	0.393	8.6
1,1,2-Trichloroethane	0.358	0.393	0.396	0.352	0.347	0.313	0.360	8.6
Tetrachloroethene	0.351	0.384	0.398	0.360	0.360	0.316	0.362	7.8
1,2-Dichloroethane-d4	0.709	0.581	0.700	0.747	0.762		0.700	10.2
Dibromofluoromethane	0.310	0.278	0.331	0.354	0.349		0.325	9.6
Toluene-d8	1.157	0.994	1.171	1.246	1.232		1.160	8.7
4-Bromofluorobenzene	0.441	0.371	0.428	0.452	0.448		0.428	7.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.