

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135148
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158
CCC	VP135152
Internal Standard/PEM	
ICV/I.BLK	VP135159
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047347.D	15 Aug 2025 08:28		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX047348.D	15 Aug 2025 09:18	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX047349.D	15 Aug 2025 09:40	LR-58,81	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX047350.D	15 Aug 2025 10:01		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX047351.D	15 Aug 2025 10:22		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX047352.D	15 Aug 2025 10:43		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX047353.D	15 Aug 2025 11:05		JC/MD	Ok,M
8	IBLK	IBLK	VX047354.D	15 Aug 2025 11:26		JC/MD	Ok
9	VSTDIC001	VSTDIC001	VX047355.D	15 Aug 2025 12:08	Not used	JC/MD	Not Ok
10	VSTDIC001	VSTDIC001	VX047356.D	15 Aug 2025 12:30		JC/MD	Ok,M
11	VSTDICV050	ICVVX081525	VX047357.D	15 Aug 2025 13:11		JC/MD	Ok,M
12	VX0815WBL01	VX0815WBL01	VX047358.D	15 Aug 2025 13:39		JC/MD	Ok
13	VX0815MBL01	VX0815MBL01	VX047359.D	15 Aug 2025 14:00		JC/MD	Ok
14	VX0815WBS01	VX0815WBS01	VX047360.D	15 Aug 2025 14:21		JC/MD	Ok,M
15	VX0815WBSD01	VX0815WBSD01	VX047361.D	15 Aug 2025 14:49		JC/MD	Ok,M
16	Q2880-01	STORAGE-BLANK-SO	VX047362.D	15 Aug 2025 15:10	vial A pH<2	JC/MD	Ok
17	Q2880-02	STORAGE-BLANK-WA	VX047363.D	15 Aug 2025 15:32	vial A pH<2	JC/MD	Ok
18	Q2880-03	STORAGE-BLANK-WA	VX047364.D	15 Aug 2025 15:53	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135148		
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158		
CCC	VP135152		
Internal Standard/PEM			
ICV/I.BLK	VP135159		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2880-04	STORAGE-BLANK-SAL	VX047365.D	15 Aug 2025 16:15	vial A pH<2	JC/MD	Ok
20	Q2886-01	BP-TT182D-TB-202508	VX047366.D	15 Aug 2025 16:36	vial A pH<2	JC/MD	Ok
21	Q2886-02	BP-TT182D-GW-202508	VX047367.D	15 Aug 2025 16:58	vial A pH<2 Surrogate Fail	JC/MD	ReRun
22	Q2837-01	TW-WTS-13	VX047368.D	15 Aug 2025 17:19	vial B pH<2	JC/MD	Ok,M
23	VSTDCCC050	VSTDCCC050EC	VX047369.D	15 Aug 2025 17:41		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081925

Review By	John Carlone	Review On	8/20/2025 8:46:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/20/2025 8:58:28 AM
SubDirectory	VX081925	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135170
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135171,VP135172

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047399.D	19 Aug 2025 08:38		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047400.D	19 Aug 2025 09:37	pH#Lot#V12668	JC/MD	Ok,M
3	VX0819MBL01	VX0819MBL01	VX047401.D	19 Aug 2025 10:05		JC/MD	Ok
4	VX0819WBL01	VX0819WBL01	VX047402.D	19 Aug 2025 10:55		JC/MD	Ok
5	VX0819WBS01	VX0819WBS01	VX047403.D	19 Aug 2025 11:55		JC/MD	Ok,M
6	Q2815-11	TW-22M-W	VX047404.D	19 Aug 2025 12:16	vial B pH<2	JC/MD	Ok
7	VX0819WBSD01	VX0819WBSD01	VX047405.D	19 Aug 2025 12:37		JC/MD	Ok,M
8	Q2815-12	TW-22M-S	VX047406.D	19 Aug 2025 12:59	vial B pH<2	JC/MD	Ok
9	Q2815-16RE	TW-17M-SRE	VX047407.D	19 Aug 2025 13:20	vial B pH<2 Surrogate Fail	JC/MD	Confirms
10	Q2815-24	TB	VX047408.D	19 Aug 2025 13:41	vial A pH<2 TB	JC/MD	Ok
11	Q2815-26	FB	VX047409.D	19 Aug 2025 14:07	vial A pH<2 FB;Surrogate Fail	JC/MD	ReRun
12	Q2841-01	MW-06-6.5-081225	VX047410.D	19 Aug 2025 14:29	vial A pH<2	JC/MD	Dilution
13	Q2841-01DL	MW-06-6.5-081225DL	VX047411.D	19 Aug 2025 14:50	vial A pH<2	JC/MD	Ok
14	Q2815-20	TW-11M-W	VX047412.D	19 Aug 2025 15:11	vial A pH<2	JC/MD	Ok
15	Q2815-21	TW-11M-E	VX047413.D	19 Aug 2025 15:32	vial A pH<2 Surrogate Fail	JC/MD	ReRun
16	Q2815-22	TW-11M-S	VX047414.D	19 Aug 2025 15:53	vial A pH<2 Surrogate Fail	JC/MD	ReRun
17	Q2815-23	TW-11M-N	VX047415.D	19 Aug 2025 16:14	vial A pH<2	JC/MD	Ok
18	Q2869-07	EB01-081325	VX047416.D	19 Aug 2025 16:36	vial A pH<2 EB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081925

Review By	John Carlone	Review On	8/20/2025 8:46:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/20/2025 8:58:28 AM
SubDirectory	VX081925	HP Acquire Method	HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135170		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135171,VP135172		

19	Q2869-10	EB02-081325	VX047417.D	19 Aug 2025 16:57	vial A pH<2 EB	JC/MD	Ok
20	Q2869-11	TB01-081325	VX047418.D	19 Aug 2025 17:18	vial A pH<2 TB	JC/MD	Ok
21	Q2815-18	TW-84SB-W	VX047419.D	19 Aug 2025 17:39	vial A pH<2	JC/MD	Ok
22	Q2815-19	DUP	VX047420.D	19 Aug 2025 18:00	vial A pH<2	JC/MD	Ok
23	Q2869-01	MW-19B-71.5-081325	VX047421.D	19 Aug 2025 18:22	vial A pH<2	JC/MD	Ok
24	VX0819MBS01	VX0819MBS01	VX047422.D	19 Aug 2025 18:43		JC/MD	Ok,M
25	Q2844-09DL	MOO-25-0226DL	VX047423.D	19 Aug 2025 19:04		JC/MD	Ok
26	Q2884-03DL	302DL	VX047424.D	19 Aug 2025 19:25		JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX047425.D	19 Aug 2025 19:47		JC/MD	Ok,M

M : Manual Integration